

Listing and evaluation of the results

68104: Maciej Przybylski

Laboratorium Mikrobiologii - DSK, Zakład Diagnostyki Laboratoryjnej i Immunologii Klinicznej Wieku

Survey of 28 April 2020

Adviser:

Prof. Dr. rer. nat. Heinz Zeichhardt
 Institut für Qualitätssicherung in der
 Virusdiagnostik - IQVD
 Potsdamer Chaussee 80
 14129 Berlin

Tel.: +49 30 8105 4300
 Fax: +49 30 8105 4303
 Mail: Heinz.Zeichhardt@iqvd.de

340

Virus Genome Detection - Coronaviruses

SARS-CoV-2 (RNA) - qualitative

Name of test	Unit	Method	Manufacturer				+
LightMix Modular Wuhan CoV RdRP-gene		PCR / NAT (395)	TIB MOLBIOL (TM)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria	
	340059	positive (3)					
	340060	below detection limit/negative (1)					
	340061	positive (3)	positive (3)	SV		+	
	340062	below detection limit/negative (1)	below detection limit/negative (1)	SV		+	
	340063	positive (3)	positive (3)	SV		+	
	340064	positive (3)	below detection limit/negative (1)	SV		+	
	340065	below detection limit/negative (1)					
Primer / Gene region	Sample	Your value					
	340059	RdRP gene					
	340060	RdRP gene					
	340061	RdRP gene					
	340062	RdRP gene					
	340063	RdRP gene					
	340064	RdRP gene					
	340065	RdRP gene					
Name of test	Unit	Method	Manufacturer				
LightMix Modular SARS and Wuhan CoV E-gene		PCR / NAT (395)	TIB MOLBIOL (TM)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria	
	340059	positive (3)					
	340060	below detection limit/negative (1)					
	340061	positive (3)	positive (3)	SV		+	
	340062	below detection limit/negative (1)	below detection limit/negative (1)	SV		+	
	340063	positive (3)	positive (3)	SV		+	
	340064	positive (3)	below detection limit/negative (1)	SV		+	
	340065	below detection limit/negative (1)					
Primer / Gene region	Sample	Your value					
	340059	E gene					
	340060	E gene					
	340061	E gene					
	340062	E gene					
	340063	E gene					
	340064	E gene					
	340065	E gene					